

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 16 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130124 ✓
Name		NICHOLAS YLIENTEMI	ID# 3429 Date 6/10/10
A	Agency		METLAKATLA POLICE DEPARTMENT Phone # 886-4011
Instrument Location		METLAKATLA POLICE DEPARTMENT ✓	
Alco S/N		X172942 ✓	Target Value .084 High Pressure 250 psi
B	Alco Test Values		1 <sup>st</sup> Alco .086 ✓ 2 <sup>nd</sup> Alco .086 ✓
Signature		Nicholas Ylimentemi ✓	
		COB 6/18/10	
(OVER)			

I, Nita J. Bolz, after being first duly sworn, depose and state as follows (Do Not write in the area below)


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 21<sup>st</sup> day of July, 2010.

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office

(Notary Seal Stamp)



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130124 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130124 ✓

JUNE 10, 2010 ✓

OPERATOR'S NAME:

VLININI/NICHOLAS/L

OPERATOR'S NUMBER: 3429

SUBJECT'S LAST NAME:

VER

SUBJECT'S FIRST NAME/MI :

OF

O.L. #: CAL

DEPT/AGENCY: MET1

CASE/REPORT: 0000

TEST TYPE: V

ALCO TARGET VALUE: .004 ✓

ALCO S/N: X172942 ✓

--- BREATH ANALYSIS ---

.004 ADJUSTED FOR 30.19 in ✓

ALCO TARGET .004 09:14

BLANK TEST .000 09:15

INTERNAL STANDARD VERIFIED 09:15

ALCO TV 30.19 in .006 ✓ 09:15

BLANK TEST .000 09:16

SUBJECT SAMPLE .000 09:17

BLANK TEST .000 09:17

ALCO TV 30.19 in .006 ✓ 09:18

BLANK TEST .000 09:19

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130124 ✓

JUNE 10, 2010 ✓

TIME 09:09 ✓

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 42c

BAROMETER: 30.19 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~